

University of Groningen

Proton Radiography to Improve Proton Radiotherapy

Biegun, Aleksandra; Takatsu, Jun; van Goethem, Marc-Jan; van der Graaf, Emiel; van Beuzekom, Martin; Visser, Jan; Brandenburg, Sijtze

Published in:
Acta Physica Polonica B

DOI:
[10.5506/APhysPolB.47.329](https://doi.org/10.5506/APhysPolB.47.329)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
Publisher's PDF, also known as Version of record

Publication date:
2016

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Biegun, A., Takatsu, J., van Goethem, M.-J., van der Graaf, E., van Beuzekom, M., Visser, J., & Brandenburg, S. (2016). Proton Radiography to Improve Proton Radiotherapy: Simulation Study at Different Proton Beam Energies. *Acta Physica Polonica B*, 47(2), 329-334.
<https://doi.org/10.5506/APhysPolB.47.329>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

PROTON RADIOGRAPHY TO IMPROVE PROTON RADIOTHERAPY: SIMULATION STUDY AT DIFFERENT PROTON BEAM ENERGIES

A.K. BIEGUN^{a†}, J. TAKATSU^b, M.-J. VAN GOETHEM^c
E.R. VAN DER GRAAF^a, M. VAN BEUZEKOM^d, J. VISSER^d,
S. BRANDENBURG^a

^aKVI-Center for Advanced Radiation Technology
University of Groningen, The Netherlands

^bDepartment of Radiation Oncology, Graduate School of Medicine
Osaka University, Japan

^cDepartment of Radiation Oncology, University Medical Center Groningen
University of Groningen, The Netherlands

^dNational Institute for Subatomic Physics (Nikhef), Amsterdam, The Netherlands

(Received October 21, 2015)

To improve the quality of cancer treatment with protons, a translation of X-ray Computed Tomography (CT) images into a map of the proton stopping powers needs to be more accurate. Proton stopping powers determined from CT images have systematic uncertainties in the calculated proton range in a patient of typically 3–4% and even up to 10% in a region containing bone. As a consequence, part of a tumor may receive no dose, or a very high dose can be delivered in healthy tissues and organs at risks (*e.g.* brain stem). A transmission radiograph of high-energy protons measuring proton stopping powers directly will allow to reduce these uncertainties, and thus improve the quality of treatment. The best way to obtain a sufficiently accurate radiograph is by tracking individual protons traversing the phantom (patient). In our simulations, we have used an ideal position sensitive detectors measuring a single proton before and after a phantom, while the residual energy of a proton was detected by a BaF₂ crystal. To obtain transmission radiographs, different phantom materials have been irradiated with a 3×3 cm² scattered proton beam, with various beam energies. The simulations were done using the Geant4 simulation package. In this study, we focus on the simulations of the energy loss radiographs for various proton beam energies that are clinically available in proton radiotherapy.

DOI:10.5506/APhysPolB.47.329

[†] Corresponding author: a.k.biegun@rug.nl

1. Introduction

Proton radiography is one of the novel imaging modalities that has a big potential to be used in proton radiotherapy as a tool for a patient positioning and as an alternative imaging tool in proton treatment. It can reduce the systematic uncertainties in the currently calculated proton range [1–9]. It delivers direct information about proton stopping powers of different materials in an object through which the proton beam has passed. The image quality of a proton radiograph is reduced by the multiple Coulomb scattering and energy loss processes of protons in matter. In our study, we applied a cut on the proton scattering angle that optimizes the quality of the reconstructed energy loss radiographs in terms of contrast and statistical accuracy.

2. Proton radiography setup

The setup that was used to simulate the energy loss and scattering angle radiographs is presented in Fig. 1. Two ideal (100% efficiency) position sensitive detectors with a size of $3 \times 3 \text{ cm}^2$ and $10 \times 10 \text{ cm}^2$ are placed before and after the phantom, respectively. They measured a position of an individual proton, as it is shown to be the best way to obtain images with the highest accuracy [4, 6, 10].




Fig. 1. (Colour on-line) A proton radiography setup used in the Geant4 simulations. Two ideal position sensitive detectors (blue squares) and an energy detector (yellow cylinder) are shown. A scattered proton beam and a phantom containing three inserts are also presented.

A BaF_2 energy detector with the diameter of 15 cm and a length of 15 cm was placed after the second position detector to measure the residual energy of a proton. A phantom with a size of 2.5 cm diameter and a length of 2.5 cm was located between position detectors. It was made of CT solid water (Gammex 457, $\rho = 1.015 \text{ g/cm}^3$) and filled with PMMA ($\rho = 1.19 \text{ g/cm}^3$), and tissue-like materials: adipose (Gammex 453, $\rho = 0.92 \text{ g/cm}^3$) and corti-

cal bone (Gammex 450, $\rho = 1.82 \text{ g/cm}^3$) [11]. A scattered proton beam with a size of $3 \times 3 \text{ cm}^2$ and different proton beam energies, $E_p = 70 \text{ MeV}$ up to $E_p = 230 \text{ MeV}$ (with a step of 20 MeV) were used to irradiate the phantom.

3. Various proton beam energies and proton scattering angle cut

To see the effect of the proton beam energy, E_p , on the energy radiograph of the phantom, Geant4 simulations [12] with various proton beam energies were performed. Proton beam energies were selected in the range available in proton radiotherapy (*i.e.* from $E_p = 70 \text{ MeV}$ up to $E_p = 230 \text{ MeV}$). In this paper, we show results for four of the selected energies: $E_p = 90, 150, 190$ and 230 MeV . Different maximum scattering angles of the proton, such as: $17.4, 8.7, 5.2$ and 1.7 mrad were applied to improve the image quality. The results for a proton beam energy of 150 MeV showed that the best trade-off between the image quality and efficiency was obtained for a proton maximum scattering angle of 5.2 mrad . For this cut, nearly 50% of the protons were used, while the image quality is almost not affected [13]. For the maximum proton scattering angle of 1.7 mrad , the image quality was the best, but a very high percentage of rejected protons (90.7%) made the cut highly inefficient. This trend is also observed for other proton beam energies, shown in Table I. For increasing proton beam energy and a selected angular cut, the number of rejected protons decreases, and thus more protons are considered to build a proton radiograph. Number of rejected protons at angular cut of 1.7 mrad is very high, up to 82% at $E_p = 230 \text{ MeV}$ (Table I), thus the cut remains inefficient. Therefore, in this paper, the energy loss radiographs for various proton beam energies, E_p , are shown for a proton scattering angle cut of 5.2 mrad .

TABLE I

Number of rejected events for maximum proton scattering angles.

$E_p \text{ [MeV]}$	Maximum proton scattering angle	
	$< 1.7 \text{ [mrad]}$	$< 5.2 \text{ [mrad]}$
90	97.6%	82.2%
150	90.7%	57.7%
190	86.3%	50.3%
230	81.8%	44.9%

3.1. Energy radiographs for $E_p = 90, 150, 190$ and 230 MeV

Energy radiographs at four selected proton beam energies (lower, middle and the highest available in clinics), with the angular cut of 5.2 mrad, are depicted in Fig. 2. At all four proton beam energies and the applied angular cut, the sharp edges between materials are visible. To determine the sharpness of the boundaries between materials in the phantom, projections through the phantom (in x - and y -directions) were evaluated.




Fig. 2. Proton radiographs at four proton beam energies of $E_p = 90$ MeV (top left), $E_p = 150$ MeV (top right), $E_p = 190$ MeV (bottom left) and $E_p = 230$ MeV (bottom right). A selection of protons with the maximum scattering angle of 5.2 mrad was applied. The color scale is adjusted for better visibility of the images.

3.2. Projections for different proton beam energies

Projections in x -direction at $y = 0.5$ cm at proton beam energies $E_p = 90, 150, 190$ and 230 MeV are shown in Fig. 3 (left). The projections were done for a single bin (bin width: 0.3 mm) with CT solid water and PMMA of proton radiographs in Fig. 2.

In all projections (also in the y -direction through CT solid water and cortical bone, not shown in this paper), the sharpness of the edges between materials for presented energies are comparable, as can be particularly seen




Fig. 3. (Colour on-line) (Left) Projections in x -direction for $y = 0.5$ cm at four proton beam energies: $E_p = 90$ MeV (black solid line), 150 MeV (red dashed line), 190 MeV (blue dashed line) and 230 MeV (black dashed line). (Right) Scaled projections from Fig. 3, left to proton beam energy of $E_p = 150$ MeV, and zoomed between 14 MeV and 18 MeV of the energy loss.

in the scaled histograms in Fig. 3 (right). After scaling, no differences in shapes and fall-offs between materials at the four demonstrated proton beam energies are noticeable. Therefore, the angular cut of 5.2 mrad can be applied for determining edges between materials independently of the proton beam energy used for a phantom irradiation.

The sharpness of the boundaries between materials characterized by the so-called *Delta* parameter are calculated as a difference in position taken at 90% and 10% of the slope between phantom materials, such as CT solid water and either PMMA, cortical bone or adipose (Fig. 4, left). For presented proton beam energies: $E_p = 90, 150, 190$ and 230 MeV, the *Delta* parameter is comparable and lower than 1.8 mm (Fig. 4, right).




Fig. 4. (Left) Definition of the *Delta* parameter. (Right) *Delta* parameter for different proton beam energies with the angular cut of 5.2 mrad.

4. Summary

In this paper, we analyze proton energy loss radiographs for various proton beam energies that are available in proton radiotherapy. The best energy loss radiograph with sufficient number of accepted protons at various proton beam energies was obtained for a proton scattering angle cut of 5.2 mrad. Therefore, this angular cut was applied to obtain energy loss radiographs at proton beam energies of $E_p = 90, 150, 190$ and 230 MeV. After scaling the images, it can be seen that the edges between materials in the phantom are equally sharp for different proton beam energies (Fig. 3, right), making the cut very efficient.

Further study with a more complex phantom containing more tissue-equivalent materials and more materials inserted on the beam path, which simulates more realistic patient geometry, is being performed.

We would like to thank the staff of the Millipede Cluster of the University of Groningen in The Netherlands, where the Monte Carlo simulations were performed (www.rug.nl/cit/hpcv/faciliteiten/HPCCluster). This work was partially supported by the Japan Society for Promotion of Science Core-to-Core Program (number 23003).

REFERENCES

- [1] U. Schneider, E. Pedroni, *Med. Phys.* **22**, 353 (1995).
- [2] U. Schneider, E. Pedroni, A. Lomax, *Phys. Med. Biol.* **41**, 111 (1996).
- [3] W. Schneider, T. Bortfeld, W. Schlegel, *Phys. Med. Biol.* **45**, 459 (2000).
- [4] G. Cirrone *et al.*, *Nucl. Instrum. Methods A* **576**, 194 (2007).
- [5] H. Paganetti, *Phys. Med. Biol.* **57**, R99 (2012).
- [6] T. Plautz *et al.*, *IEEE T. Med. Imaging* **33**, 875 (2014).
- [7] G. Landry *et al.*, *Phys. Med. Biol.* **58**, 5029 (2013).
- [8] J. Schuemann *et al.*, *Phys. Med. Biol.* **59**, 4007 (2014).
- [9] A-C. Knopf, A. Lomax, *Phys. Med. Biol.* **58**, R131 (2013).
- [10] V. Sipala *et al.*, *JINST* **8**, C02021 (2013).
- [11] <http://www.gammex.com/>
- [12] S. Agostinelli *et al.*, *Nucl. Instrum. Methods A* **506**, 250 (2003).
- [13] J. Takatsu *et al.*, *JINST* **11**, C01004 (2016).